

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

MAY 27 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

|  |   |   |
|--|---|---|
| Supervisor/Operator Performing the Verification Procedure: |   | DataMaster cdm S/N <u>130130</u> ✓                  |
| Name <u>JEREMY CONKLIN</u>                                 |   | ID# <u>6715</u> Date <u>05.21.09</u> ✓              |
| A  | Agency <u>VALDEZ POLICE DEPARTMENT</u>                        | Phone # <u>835-4560</u>                             |
| Instrument Location <u>VALDEZ JAIL</u> ✓                   |   |   |
| Alco S/N <u>70108</u> ✓                                    |   | Target Value <u>.083</u> ✓ High Pressure <u>500</u> |
| B  | Alco Test Value Average <u>.081</u> ✓<br>1 <sup>st</sup> Alco | <u>.080</u> ✓<br>2 <sup>nd</sup> Alco               |
| Signature <u>[Signature]</u> 271 ✓                         |   |   |
| (OVER)   |   | <u>BMB</u><br><u>5/27/09</u>                        |


Nita J. Bolz, after being first duly sworn, depose and state as follows:  
(Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 27<sup>th</sup> day of May, 2009.

[Signature]  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130130

MAY 21, 2009

OPERATOR'S NAME:  
CONNELL

OPERATOR'S NUMBER: 6715

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME (1):

UDC

O.L. #1 UDC

DEPT AGENCY: UDZ1

TEST REPORT: UDC

TEST TARGET

ALCO TARGET VALUE: .083

ALCO S/N: 70108

--- BREATH ANALYSIS ---

.083 ANALYZED AIR 30.43 in

ALCO TEST .004 01:54

BLANK TEST .000 01:55

INTERNAL STANDARD VERIFY 01:56

ALCO TV 30.43 in .001 01:56

BLANK TEST .000 01:57

SUBJECT SAMPLE .000 01:58

BLANK TEST .000 01:59

ALCO TEST .000 01:59

BLANK TEST .000 01:59

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130130

MAY 21, 2009

TIME 01:48

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 6/2/00/01

HEATERS

SAMPLE TUBE: 49c

PROB TUBE: 49c

BAROMETER: 30.45 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

ALCO S/N: 70108  
HIZKLMNOPQRSTU VWXYZ 1234567890  
PQRSTUVWXYZ 1234567890